

## EC – DECLARATION OF CONFORMITY

The product named hereinafter was developed, designed, and manufactured in compliance with the relevant, fundamental safety and health requirements of the listed EC directives and norms. In the event of modifications that were not authorised by us or if the product is used in a manner

that is not in line with the intended purpose, this declaration will be rendered void.

Product name: Laboratory centrifuge

Product type: Sigma 3-16KL

Order number: 10360, 10361, 10362, 91027

Directives: 2006/42/EC Machinery Directive

2014/35/EU Low Voltage Directive

2014/30/EU EMC Directive (EU) 2015/863 RoHS Directive

Normes: EN 61010-2-020:2017

EN 61010-2-011:2017 EN IEC 61000-3-2:2019

EN 61000-3-3:2013 EN 61326-1:2013

Sigma Laborzentrifugen GmbH

An der Unteren Söse 50 37520 Osterode

Germany

Authorised representative

for CE matters: Eckhard Tödteberg

Osterode, 07/04/2020

Michael Sander

General Manager

CE\_3-16KL\_2020-04-07\_en Page 1 / 1



## EC – DECLARATION OF CONFORMITY

The product named hereinafter was developed, designed, and manufactured in compliance with the relevant, fundamental safety and health requirements of the listed EC directives and norms. In the event of modifications that were not authorised by us or if the product is used in a manner

that is not in line with the intended purpose, this declaration will be rendered void.

Product name:

Laboratory centrifuge

Product type:

Sigma 3-16KL IVD

Order number:

10363, 10364

Directives:

98/79/EC In vitro Diagnostica (EU) 2015/863 RoHS Directive

Normes:

EN 61010-2-101:2017
EN 61010-2-020:2017

EN IEC 61000-3-2:2019 EN 61000-3-3:2013 EN 61326-1:2013

EN 61010-2-011:2017

Sigma Laborzentrifugen GmbH

An der Unteren Söse 50 37520 Osterode

Germany

Authorised representative

for CE matters: Eckhard Tödteberg

Osterode, 07/04/2020

Michael Sander

General Manager

CE\_3-16KL\_IVD\_2020-04-07\_en Page 1 / 1