

Eppendorf Certificate

Certificate of Quality

Eppendorf Centrifuges MiniSpin®, MiniSpin® plus, 5415 D, 5415 R, 5417 R, 5418, 5418 R, 5424, 5424 R, 5425, 5425 R, 5427 R, 5430, 5430 R, 5702, 5702 R, 5702 RH, 5804, 5804 R, 5810, 5810 R, 5910 R, and 5920 R

The Eppendorf Quality Management System is certified in accordance with the international standards ISO 9001, ISO 13485, and ISO 14001.

Eppendorf Centrifuges are tested in accordance with current manufacturing specifications and safety requirements.

Testing with respect to compliance within the defined limitations and tolerances includes assessment of

- rotational speed,
- run time,
- cooling performance, if applicable.

For Operational Qualification by certified service personnel, documentation including the defined limitations and tolerances is available and can be provided upon request.

Hamburg, September 2020

Eppendorf Certificate

Certificate of Quality - Addendum

Eppendorf Centrifuges MiniSpin®, MiniSpin® plus, 5415 D, 5415 R, 5417 R, 5418, 5418 R, 5424, 5424 R, 5425, 5425 R, 5427 R, 5430, 5430 R, 5702, 5702 R, 5702 RH, 5804, 5804 R, 5810, 5810 R, 5910 R, and 5920 R

Limitations and tolerances

Eppendorf® Centrifuge	Rotor	Rotational Speed	Temperature	Run Time
MiniSpin®	F-45-12-11	13,400 rpm ±402 rpm	ΔT ≤ 14 K	300 s ± 2 s
MiniSpin® plus	F-45-12-11	14,500 rpm ±435 rpm	ΔT ≤ 14 K	300 s ± 2 s
5418	FA-45-18-11	14,000 rpm ±100 rpm	ΔT ≤ 12 K	300 s ± 2 s
5418 R	FA-45-18-11	14,000 rpm ±100 rpm	4 °C ±2 K	300 s ± 2 s
5424 (100 V / 230 V)	FA-45-24-11	14,680 rpm ±100 rpm	ΔT ≤ 18 K	300 s ± 2 s
5424 (120 V)	FA-45-24-11	15,000 rpm ±100 rpm	ΔT ≤ 18 K	300 s ± 2 s
5424 R	FA-45-24-11	14,680 rpm ±100 rpm	4 °C ±2 K	300 s ± 2 s
5425	FA-24x2	15,060 rpm ±40 rpm	ΔT ≤ 17 K	300 s ± 3 s
5425 R	FA-24*2	15,060 rpm ±40 rpm	4 °C ±2 K	300 s ± 2 s
5427 R	F-45-48-11	14,000 rpm ±50 rpm	4 °C ±2 K	300 s ± 2 s
5427 R	F-45-30-11	14,000 rpm ±50 rpm	4 °C ±2 K	300 s ± 2 s
5430	FA-45-30-11	14,000 rpm ±100 rpm	ΔT ≤ 14 K	300 s ± 2 s
5430 R	FA-45-30-11	14,000 rpm ±100 rpm	4 °C ±2 K	300 s ± 2 s
5702	F-35-30-17	4,400 rpm ±44 rpm	ΔT ≤ 12 K	300 s ± 2 s
5702 R	A-4-38	4,400 rpm ±44 rpm	4 °C ±2 K @ 4,400 rpm	300 s ± 2 s
5702 RH	A-4-38	4,400 rpm ±44 rpm	4 °C ±2 K @ 4,400 rpm, 37 °C ±2 K @ 2,200 rpm	300 s ± 2 s
5804	A-4-44	5,000 rpm ±50 rpm	ΔT ≤ 13 K	300 s ± 2 s
5804 R	A-4-44	5,000 rpm ±50 rpm	7 °C ±2 K	300 s ± 2 s
5810	A-4-62	4,000 rpm ±40 rpm	ΔT ≤ 13 K	300 s ± 2 s
5810	F-34-6-38	12,000 rpm ±40 rpm	ΔT ≤ 16 K	300 s ± 2 s
5810	A-4-81	4,000 rpm ±40 rpm	ΔT ≤ 13 K	300 s ± 2 s
5810 R	A-4-62	4,000 rpm ±40 rpm	6 °C ±2 K	300 s ± 2 s
5810 R	F-34-6-38	12,000 rpm ±40 rpm	13 °C ±2 K	300 s ± 2 s
5810 R	A-4-81	4,000 rpm ±40 rpm	8 °C ±2 K	300 s ± 2 s
5910 R	S4xUniversal	14,000 rpm ±40 rpm	4 °C ±2 K	300 s ± 3 s
5920 R	S4xUniversal-Large	14,000 rpm ±40 rpm	4 °C ±2 K	300 s ± 3 s

Dr. Reza Hashemi
Product Life Cycle Manager Centrifugation

Monika Schneider
Head of Global Quality Management & Regulatory Affairs

Your local distributor: www.eppendorf.com/contact
Eppendorf AG · Barkhausenweg 1 · 22339 Hamburg · Germany
E-mail: eppendorf@eppendorf.com

ISO 9001
Certified

ISO 13485
Certified

ISO 14001
Certified