

EC – DECLARATION OF CONFORMITY

The product named hereinafter was developed, designed, and manufactured in compliance with the relevant, fundamental safety and health requirements of the listed EC directives and norms.

In the event of modifications that were not authorised by us or if the product is used in a manner that is not in line with the intended purpose, this declaration will be rendered void.

Product name: Laboratory centrifuge

Product type: Sigma 4-5L

Order number: 10403, 10404, 10405

Directives:

| | |
|---------------|-----------------------|
| 2006/42/EC | Machinery Directive |
| 2014/35/EU | Low Voltage Directive |
| 2014/30/EU | EMC Directive |
| (EU) 2015/863 | RoHS Directive |

Normes:

- EN 61010-2-020:2017
- EN IEC 61000-3-2:2019
- EN 61000-3-3:2013
- EN 61326-1:2013

Sigma Laborzentrifugen GmbH

An der Unteren Söse 50
37520 Osterode
Germany

Authorised representative
for CE matters:
Eckhard Tödteberg

Osterode, 07/04/2020



General Manager

EC – DECLARATION OF CONFORMITY

The product named hereinafter was developed, designed, and manufactured in compliance with the relevant, fundamental safety and health requirements of the listed EC directives and norms. In the event of modifications that were not authorised by us or if the product is used in a manner that is not in line with the intended purpose, this declaration will be rendered void.

Product name: Laboratory centrifuge

Product type: Sigma 4-5L IVD

Order number: 10406, 10407, 10408

Directives: 98/79/EC In vitro Diagnostica
(EU) 2015/863 RoHS Directive

Normes: EN 61010-2-101:2017
EN 61010-2-020:2017
EN IEC 61000-3-2:2019
EN 61000-3-3:2013
EN 61326-1:2013

Sigma Laborzentrifugen GmbH

An der Unteren Söse 50
37520 Osterode
Germany

Authorised representative
for CE matters:
Eckhard Tödteberg

Osterode, 07/04/2020



General Manager