# AVANTI J-15 SERIES


THE RELIABILITY AND PERFORMANCE YOU TRUST





## FROM GENERAL CELL CULTURE PREPARATIONS TO PURIFIED END PRODUCT APPLICATIONS

THE NEW AVANTI J-15 SERIES IS PART OF A NEW FAMILY OF LIFE SCIENCE EQUIPMENT THAT PROVIDES THE EXCEPTIONAL PERFORMANCE YOU EXPECT FROM BECKMAN COULTER.



- 9 Berggren, W. Travis, Margaret Eutz, and Veronica Modesto. General spiniection protocol. Stembook (2012
- ¥ Fuss, Ivan J., et al. Isolation of whole mononuclear cells from peripheral blood and cord blood. John Wiley & Sons, Inc., 2009
- \* BioCertified has been tested and validated to demonstrate containment of microbial aerosols by an independent third party facility (Health Protection Agency, Porton Down, UK or USAMRIID, Ft. Detrick, MD, USA).

## ENHANCED CONTROL OF SAMPLE, TIME AND WORKFLOW

## **ULTRA**Harmonic

The Ultra Harmonic Technology optimizes sample separation by reducing the net changes of forces to the sample, maximizing the amount of specimen presence in the pellet, and reducing the presence of contaminants or unpelleted materials in the supernatant.

It also optimizes time by implementing faster acceleration and deceleration profiles without compromising on sample quality.



### **MULTI-DIMENSIONAL DIFFERENTIATION & FEATURES**



#### Early Imbalance Detection Within 30 seconds

#### Increases confidence on experimental runs

We've optimized our sensitivity to detect rotor imbalance as close to the beginning of the run as possible. That means the user can set the experiment and walk away with confidence. This saves time and protects the sample.



#### **True Imbalance Detection**

#### Reduces false diagnostics

Our improved hardware and software accurately respond to true imbalance without triggering nuisance imbalance diagnostics. This feature reduces the hassle and downtime during the workflow. System allows up to 12 gram imbalance tolerance\*



#### **Lower Lift Over Height**

#### Easy rotor swapping

Designed with the end-user in mind, the J-15 has a 13% lower lift over height when compared to the Allegra X-series centrifuges. This allows for easier rotor swapping for all user types.

## **ROTORS AVAILABLE**

#### **Swinging Bucket**



#### JS-4.750

- Max RPM: 4,739 RPM
- Max RCF: 5,250 x g
- 3 liter Maximum Capacity
- BioSafe options
- Sample containers supported
  - Flasks
  - Tube range: 3 50mL
  - Bottles: 230 conical, 250, 500, 750mL
  - Aerosolve Canisters



#### JS-4.750u

- Max RPM: 4,450 RPM
- Max RCF: 4,060 x g
- BioSafe options
- Plate Capacity
  - 16 single plates
  - 4 deep well

#### **Fixed Angle**



#### **JA-10.100**

- Max RPM: 10,200 RPM
- Max RCF: 11,420 x g
- 600mL Maximum Capacity
- BioSafe Certification
- Sample containers supported
  - 1.5, 10, 15, 50, 100mL tubes

Item No.	Item Description	
B77580	Rotor Assembly, JS-4.750	
B77584	Rotor Assembly, JA-10.100	
B83980	Rotor Assembly, JS-4.750 Microplate Carriers	

<sup>\*</sup> Using the JS-4.750 rotor

## **INSTRUMENT AND ROTOR PACKAGES**

#### **Swinging Bucket**

#### Avanti J-15 series swinging bucket packages

Includes benchtop centrifuge and JS-4.750 (4 x 750 mL) swinging bucket rotor.

Bottles are NOT included and can be ordered separately

Part Number	Description
C19399	Avanti J-15, IVD, Ventilated, 100-120 V, 50/60 Hz, Swinging Bucket Package
C19398	Avanti J-15R, IVD, Refrigerated, 120 V, 60 Hz, Swinging Bucket Package
C19400	Avanti J-15, Ventilated, 100-120 V, 50/60 Hz, Swinging Bucket Package
C19397	Avanti J-15R, Refrigerated, 120 V, 60 Hz, Swinging Bucket Package



- 4,750 RPM/5,250 x g RCF (J-15R 200-230V)
- 4,550RPM/4,820 x g RCF (J-15/R 120V)

#### Avanti J-15 series microplate packages

Includes benchtop centrifuge and JS-4.750µ microplate rotor (yoke with microplate carriers), 4 microplate carriages

Microplates and microplate covers are NOT included and can be ordered separately.

Part Number	Description
C19405	Avanti J-15, IVD, Ventilated, 100-120 V, 50/60 Hz, Microplate Package
C19404	Avanti J-15R, IVD, Refrigerated, 120 V, 60 Hz, Microplate Package
C19406	Avanti J-15, Ventilated, 100-120 V, 50/60 Hz, Microplate Package
C19403	Avanti J-15R, Refrigerated, 120 V, 60 Hz, Microplate Package



- 4,450RPM/4,060 x g RCF (J-15R)
- 4,350RPM/3,880 x g RCF (J-15)

#### Avanti J-15 series cell culture packages

Includes benchtop centrifuge and JS-4.750 (4 x 750 mL) swinging bucket rotor, 15 mL conical tube adapters (qty 4) and 50 mL conical tube adapters (qty 4).

Bucket covers and tubes/bottles are NOT included and can be ordered separately.

Part Number	Description
C19417 Avanti J-15, IVD, Ventilated, 100-120 V, 50/60 Hz, Cell Culture Package	
C19416	Avanti J-15R, IVD, Refrigerated, 120 V, 60 Hz, Cell Culture Package
C19418	Avanti J-15, Ventilated, 100-120 V, 50/60 Hz, Cell Culture Package
C19415	Avanti J-15R, Refrigerated, 120 V, 60 Hz, Cell Culture Package



- 4,750 RPM/5,250 x g RCF (J-15R 200-230V)
- 4.550 RPM/4,820 x g RCF (J-15/R 120V)
- Adapters:
  - Holds 56 of 15 mL tubes
  - Holds 28 of 50 mL Tubes

## **INSTRUMENT AND ROTOR PACKAGES**

#### **Swinging Bucket**

#### Avanti J-15 series biosafe packages

Includes benchtop centrifuge and BioCertified\* JS-4.750 (4 x 750 mL) swinging bucket rotor, BioCertified\* aerosolve canisters (qty 4), 15 mL conical BioSafe aerosolve canister tube racks (qty 4) and 50 mL conical BioSafe aerosolve canister tube racks (qty 4). (Rotor is BioCertified\* when used with aerosolve canisters.)

Tubes/bottles are NOT included and can be ordered separately.

Part Number	Description
C19411	Avanti J-15, IVD, Ventilated, 100-120 V, 50/60 Hz, BioSafe Package
C19410	Avanti J-15R, IVD, Refrigerated, 120 V, 60 Hz, BioSafe Package
C19412	Avanti J-15, Ventilated, 100-120 V, 50/60 Hz, BioSafe Package
C19409	Avanti J-15R, Refrigerated, 120 V, 60 Hz, BioSafe Package

#### Avanti J-15 series blood sample packages

Includes benchtop centrifuge and JS-4.750 (4 x 750 mL) swinging bucket rotor, 13 mm diameter tube adapters (qty 4) and 16 mm diameter tube adapters (qty 4).

Bucket covers and tubes/bottles are NOT included and can be ordered separately.

Part Number	Description
C19423	Avanti J-15, IVD, Ventilated, 100-120 V, 50/60 Hz, Blood Sample Package
C19422	Avanti J-15R, IVD, Refrigerated, 120 V, 60 Hz, Blood Sample Package
C19424	Avanti J-15, Ventilated, 100-120 V, 50/60 Hz, Blood Sample Package
C19421	Avanti J-15R, Refrigerated, 120 V, 60 Hz, Blood Sample Package



- 4,750 RPM/5,250 x g RCF (J-15R 200-230V)
- 4,550 RPM/4820 x g RCF (J-15/R 120V)



- 120 13mm Blood Tubes
- 76 16mm Blood Tubes

#### **Fixed Angle**

#### Avanti J-15 series Fixed Angle packages

Includes benchtop centrifuge and BioCertified\* JA-10.100 (6 x 100 mL) fixed angle rotor.

Adapters, tubes/bottles are NOT included and can be ordered separately.

Part Number	Description
C19393	Avanti J-15, IVD, Ventilated, 100-120 V, 50/60 Hz, Fixed Angle Package
C19392	Avanti J-15R, IVD, Refrigerated, 120 V, 60 Hz, Fixed Angle Package
C19394	Avanti J-15, Ventilated, 100-120 V, 50/60 Hz, Fixed Angle Package
C19391	Avanti J-15R, Refrigerated, 120 V, 60 Hz, Fixed Angle Package



- 600 mL Maximum Capacity
- 10,200 RPM 11,420 x g RCF
- Biosafety Lid

## **ROTOR ACCESSORIES**

#### **Swinging Bucket**

Table 1: Additional Biosafety accessories for personal protection and sample protection

Part Number	Description
<b>392805</b> JS-4.750 Tube-Bottle Bucket Covers (Set of 2)	
393070	JS-4.750 Multi-well Plate Carrier Cover (Set of 2)
392804	JS-4.750 Replacement Tube-and-bottle bucket (set of 2)
392806	JS-4.750 Multi-well plate carrier (set of 2)
359481	JS-4.750 Aerosolve Canister (Set of 2)



Color	Nom. Tube	Nom. Tube Dia. (mm)	Max. No. of Tubes per Adapter	Max. No. Tubes in Rotor	Adapter Part No.	
Code	Vol. (mL)				Set of 2	Set of 4
blue	3 5	10 12	37	148	359469	359148
tan	3 & 5	13	30	120	359478	359157
orange	7 & 10	14	24	96	359470	359149
purple	12	16	19	76	359471	359150
green conical	15	18	14	56	359472	359151
green	15 & 20	18	14	56	359473	359152
lt. green conical	30 & 50	30	4	16	359475	359154
yellow	50	29	7	28	359474	359153
dk. blue	50	35	4	16	359476	359155



Table 3: Cell Culture Flask Adapters (EPDM)

Color Code	Flask Size (cm²)	No. Flasks per Adapter	Part No. (Qty 2)
orange	75	1	369292
green	25	2	369295



#### **Fixed Angle**

Table 4: Bottle and Tube Adapters for the JA-10.100 Rotor<sup>a</sup>

Adapter Part Number (Set of 6)	Tube or Bottle Type <sup>b</sup>	Tube Dimensions (mm)	Nominal Tube Volume (mL)
392830	round-bottom tube or bottle	29 x 108	50
392268	conical tube	30 x 115	50
392823	bottle	18 x 107	15
392270	conical tube	17 x 120	15
392824	round-bottom bottle	16 x 82	10
344497°	microfuge tube	11 x 39	1.5

- a. Unless otherwise indicated, adapters are polypropylene
- b. Observe manufacturer's recommendations for RCF and temperature limitations.
- c. 344497 fits in 392830.

## **COMPLIANCE AND TECHNICAL SUPPORT**

- At Beckman Coulter, engineering , sales, support, training and service work together to offer comprehensive and extensive customer focused products.
- Expert service engineering team strives for "Fix It Right the First Time."
- · Certifications of Compliance.






#### **Specification**

	Part Number	Centirfuge Version
IVD	B99515	AVANTI J-15R, 200-230 VAC, 50Hz
	C31547	AVANTI J-15R, 208-230 VAC, 60Hz
	B99517	AVANTI J-15R, 120 VAC, 60Hz
	C01994	AVANTI J-15
	B99514	AVANTI J-15R, 200-230 VAC, 50Hz
Non-IVD	C31546	AVANTI J-15R, 208-230 VAC, 60Hz
	B99516	AVANTI J-15R, 120 VAC, 60Hz
	C01995	AVANTI J-15

ersion	
0-230	
8-230	
) VAC,	-
0-230	
8-230	
) VAC,	_

Providing 70 years of global leadership in centrifugation, Beckman Coulter Life Sciences designs, manufactures, sells, and services a complete line of centrifuge systems. By offering unique rotors and innovative bottles, tubes and accessories. coupled with advanced centrifugation software, Beckman Coulter delivers intelligent centrifugation solutions to laboratory science.

Specification	Description	Avanti J-15	Avanti J-15R
Speed	Set Speed	200 to 10,200 RPM in 10 RPM increments	200 to 10,200 RPM in 10 RPM increments
	Set RCF	10 to 11,420 x g in 10 x g increments	10 to 11,420 x g in 10 x g increment
	Speed Display	Actual rotor speed in 10 RPM increments or actual RCF in 10 x g increments	
	Speed Accuracy	±25 RPM of Set Speed from 200 to 10,200 RPM	
	Set Time	1 minute to 99 hours and 59 minutes or continuous (Hold)	
Time	Time Display	Timed run: indicates run time remaining (HH:MM:SS) Hold run: indicates elapsed time (HH:MM:SS) Pulse run: indicates elapsed time (HH:MM:SS)	
Temperature	Set Temperature <sup>1</sup>		-10 to +40° C in 1° C increments
	Temperature Display	N/A	Chamber temperature in 1° C increments
	Temperature Accuracy		±2° C of Chamber temperature (after equilibration); applies to 4 to 25C temp range
	Over Temperature Shutdown <sup>2</sup>	> 55° C	> 55° C
Acceleration	Acceleration Profiles	10 acceleration rates, including maximum torque	
Deceleration	Deceleration Profiles	11 deceleration rates, including maximum torque and no braking	
Dimensions	Width	55.6 cm (21.9 in)	75.6 cm (29.8 in)
	Depth	74.9 cm (29.5 in)	70.3 cm (27.7 in)
	Height	36.8 cm (14.5 in)	36.8 cm (14.5 in)
Weight	Weight, not including rotor	93 kg (205 lbs)	120 kg (265 lbs)
Ventilation Clearances	Sides	30 cm (12.0 in)	7.6 cm (3.0 in)
	Rear	30 cm (12.0 in)	7.6 cm (3.0 in)
Finishes	Top Surface	Painted steel	
	Front Surface	Uncoated plastic	
	Door	Painted aluminum and plastic	
Electrical	Electrical Requirements	100V, 12A, 50/60Hz 120V, 10A, 50/60Hz	120V, 12A, 60Hz
		200-230V, 6A, 50/60Hz	200-230V, 8A, 50Hz 208-230V, 9A, 60Hz
	Electrical Supply	Class 1	
	Installation (overvoltage) Category	П	
Environmental	Noise output (1 m in front of instrument, 1.5 m above the floor with JA-10.100 rotor at 10,200 RPM)	61 dBA	58 dBA
	Ambient Temperature Range	10 to 31° C	10 to 35° C
	Humidity	80%, noncondensing	80%, noncondensing
	Refrigerant	N/A	R404A
	Maximum Heat Dissipation under steady state conditions	4095 Btu/h (1.2 kW)	120V: 4913 Btu/h (1.44 kW) 200-230V: 6551 Btu/h (1.92 kW)
	Pollution Degree <sup>3</sup>	2	
	Altitude	Up to 2,000 meters	



<sup>2</sup> If the system reaches this temperature, it will issue a diagnostic and shut down using max brake.

<sup>© 2019</sup> Beckman Coulter, Inc. All rights reserved. Beckman Coulter, the stylized logo, and the Beckman Coulter product and service marks mentioned herein are trademarks or registered trademarks of Beckman Coulter, Inc. in the United States and other countries.



<sup>3</sup> Normally, only nonconductive pollution occurs. Occasionally, however, a temporary conductivity caused by condensation must be expected.